



## Certificate of Analysis

Product Name	:	ALyS505N-0
Cat. No.	:	1020P10
Size	:	Bottle 1000 mL
Lot No.	:	
Expiration Date	:	YYYY.MM

EXAMINATIONS		SPECIFICATIONS	RESULTS
Appearance		Orange to Red, Clear Liquid	
pH *1		7.4 ± 0.2	
Osmotic Press. *2		290 ± 10 mOsm	
Endotoxin*2		< 0.25 EU/mL	
Sterility Test*2	Bacteria	Negative	
	Fungi	Negative	
Mycoplasma*2		Negative	
Cell Culture Test		PASS	

\*1 After incubate in a 5%CO<sub>2</sub> incubator.

\*2 JP

We prove that this product adapted to above specifications.

\_\_\_\_\_

QC director

\_\_\_\_\_

Acceptance Date

Note to User: Culture media are reagent for research use, do not use for clinical therapy and diagnostic procedure, and strictly prohibited an injection into a body directly.