



CELL SCIENCE & TECHNOLOGY
INSTITUTE INC.

1000-13, Araya,
Tendo, Yamagata, 994-0054, Japan
TEL : +81-23-665-5300

COA No. : A2G00P05C-008E

Certificate of Analysis

Product Name	: CiMS-BM
Cat. No.	: A2G00P05C
Size	: Bottle 500 mL
Lot No.	: 0706220E
Expiration Date	: 2023.04

EXAMINATIONS		SPECIFICATIONS	RESULTS
Appearance		Orange to red, Clear Liquid	PASS
pH *1		7.4 ± 0.2	7.4
Osmotic Press. *2		300 ± 15 mOsm	300 mOsm
Endotoxin*2		< 0.25 EU/mL	0.001 EU/mL
Sterility Test*2	Bacteria	Negative	PASS
	Fungi	Negative	PASS
Cell Culture Test		PASS	PASS

*1 After incubate at 37°C, 5%CO₂ incubator.

*2 JP

We prove that this product adapted to above specifications.



Satori Ishiya
QC director

2022.07.22

Acceptance Date

Note to User: Culture media are reagent for research use, do not use for clinical therapy and diagnostic procedure, and strictly prohibited an injection into a body directly.